

standards of purity and quality and shall be effectively segregated and otherwise manufactured in a manner that will reduce the risk of contaminating the product and other biological products. Ancillary reagents and materials accompanying the product, which are used in the performance of the test as described by the manufacturer's recommended test procedures, shall have been shown not to affect adversely the product within the prescribed dating period.

(c) *Final container.* A final container shall be sufficiently transparent to permit visual inspection of the contents for presence of particulate matter and increased turbidity. The effectiveness of the contents of a final container shall be maintained throughout its dating period.

(d) *Date of manufacture.* The date of manufacture of Hepatitis B Surface Antigen that has been iodinated with radioactive iodine (^{125}I) shall be the day of labeling the antibody with the radionuclide.

[44 FR 36382, June 22, 1979, as amended at 49 FR 1685, Jan. 13, 1984]

§ 660.43 Potency test.

To be satisfactory for release, each filling of Hepatitis B Surface Antigen shall be tested against the Reference Hepatitis B Antiserum Panel and shall be sufficiently potent to be able to detect the antibody in the appropriate sera of the reference panel by all test methods recommended by the manufacturer in the package insert.

§ 660.44 Specificity.

Each filling of the product shall be specific for Hepatitis B Surface Antigen as determined by specificity tests found acceptable to the Director, Center for Biologics Evaluation and Research.

[44 FR 36382, June 22, 1979, as amended at 49 FR 23834, June 8, 1984; 55 FR 11013, Mar. 26, 1990]

§ 660.45 Labeling.

In addition to the requirements of §§ 610.60, 610.61, and 809.10 of this chapter, the labeling shall bear the following:

(a) The "d and y" antigen subtype and the source of the product to follow immediately the proper name on both the final container label and the package label. If the product is intended to identify antibodies to the "r and w" antigen subtype, the antigen subtype designation shall include the "r and w" antigen subtype.

(b) The name of the test method(s) recommended for use of the product on the package label and on the final container label, when capable of bearing a full label (see § 610.60(a) of this chapter).

(c) A warning on the package label and on the final container label stating that the product is capable of transmitting hepatitis and should be handled accordingly.

(d) The package shall include a package insert providing (1) detailed instructions for use, (2) an adequate description of all recommended test methods, and (3) warnings as to possible hazards, including hepatitis transmitted in handling the product and any ancillary reagents and materials accompanying the product.

§ 660.46 Samples; protocols; official release.

(a) *Samples.* (1) For the purposes of this section, a sample of product not iodinated with ^{125}I means a sample from each filling of each lot packaged as for distribution, including all ancillary reagents and materials; and a sample of product iodinated with ^{125}I or unlyophilized HBsAg-coated red blood cells means a sample from each lot of diagnostic test kits in a finished package, including all ancillary reagents and materials.

(2) Unless the Director, Center for Biologics Evaluation and Research, determines that the reliability and consistency of the finished product can be assured with a smaller quantity of sample or no sample and specifically reduces or eliminates the required quantity of sample, each manufacturer shall submit the following samples to the Director, Center for Biologics Evaluation and Research (see mailing addresses in § 600.2 of this chapter), within 5 working days after the manufacturer has satisfactorily completed all tests on the samples:

§ 660.50

(i) One sample until written notification of official release is no longer required under paragraph (c)(2) of this section.

(ii) One sample of product at periodic intervals of 90 days, beginning after written notification of official release is no longer required under paragraph (c)(2) of this section. The sample submitted at the 90-day interval shall be from the first lot or filling, as applicable, released by the manufacturer, under the requirements of §610.1 of this chapter, after the end of the previous 90-day interval. The sample shall be identified as "surveillance sample" and shall include the date of manufacture.

(iii) Samples may at any time be required to be submitted to the Director, Center for Biologics Evaluation and Research, if the Director finds that continued evaluation is necessary to ensure the potency, quality, and reliability of the product.

(b) *Protocols.* For each sample submitted as required in paragraph (a)(1) of this section, the manufacturer shall send a protocol that consists of a summary of the history of manufacture of the product, including all results of each test for which test results are requested by the Director, Center for Biologics Evaluation and Research. The protocols submitted with the samples at periodic intervals as provided in paragraph (a)(2)(ii) of this section shall be identified by the manufacturer as "surveillance test results."

(c) *Official release.* (1) The manufacturer shall not distribute the product until written notification of official release is received from the Director, Center for Biologics Evaluation and Research, except as provided in paragraph (c)(2) of this section. Official release is required for at least five consecutive lots or fillings, as applicable, manufactured after licensure of the product.

(2) After written notification of official release is received from the Director, Center for Biologics Evaluation and Research, for at least five consecutive lots or fillings manufactured after licensure of the products, and after the manufacturer receives from the Director, Center for Biologics Evaluation and Research, written notification that official release is no longer required,

21 CFR Ch. I (4–1–10 Edition)

subsequent lots or fillings may be released by the manufacturer under the requirements of §610.1 of this chapter.

(3) The manufacturer shall not distribute lots or fillings, as applicable, of products that require sample submission under paragraph (a)(2)(iii) of this section until written notification of official release or notification that official release is no longer required is received from the Director, Center for Biologics Evaluation and Research.

[48 FR 20407, May 6, 1983, as amended at 49 FR 23834, June 8, 1984; 51 FR 15611, Apr. 25, 1986; 55 FR 11013 and 11014, Mar. 26, 1990; 70 FR 14985, Mar. 24, 2005]

Subpart F—Anti-Human Globulin

§ 660.50 Anti-Human Globulin.

(a) *Proper name and definition.* The proper name of this product shall be Anti-Human Globulin which shall consist of one or more antiglobulin antibodies identified in §660.55(d).

(b) *Source.* The source of this product shall be either serum from animals immunized with one or more human serum globulins or protein-rich fluids derived from stable immunoglobulin-secreting cell lines maintained either in tissue cultures or in secondary hosts.

[50 FR 5579, Feb. 11, 1985, as amended at 65 FR 77499, Dec. 12, 2000]

§ 660.51 Processing.

(a) *Processing method.* (1) The processing method shall be one that has been shown to yield consistently a specific, potent final product, free of properties that would adversely affect the product for its intended use throughout its dating period.

(2) Anti-IgG, -C3d (polyspecific) reagents and anti-IgG products may be colored green.

(3) Only that material which has been fully processed, thoroughly mixed in a single vessel, and filtered shall constitute a lot. Each lot shall be identified by a lot number.

(4) A lot may be subdivided into sublots which shall be identified by the lot number to which has been added a distinctive prefix or suffix. If lots are to be subdivided, the manufacturer